

# Final Report

## Calibration-Round-Robin

### CaRo 15

**Adolf Kühner AG**

Dinkelbergstr. 1  
CH-4127 Birsfelden  
Switzerland

Tel. +41 (0)61 319 93 93  
Fax. +41 (0)61 319 93 94  
E-mail office@kuhner.com  
Internet www.kuhner.com

This report shall not be published or reproduced other than in full.

CaRo 15	1
1. Conclusion	2
2. Explosion Indices Pmax, Kmax	5
3. Minimum Ignition Energy MIE	7
4. Troubles with chemical igniters	9
5. List of Participants	10

## 1. Conclusion

---

According to international standards (e.g. ISO 9000, GLP), test equipment must be calibrated at intervals by comparison with a standard or a calibrated testing apparatus.

This calibration also applies to the 20-l-apparatus and the 1-m<sup>3</sup>-vessel for the determination of P<sub>max</sub> and K<sub>max</sub> and the apparatus for determination of the minimum ignition energy. The test procedure is an important part of this calibration. A general check at the component level is incomplete and hence inadmissible.

Unfortunately there are neither internationally recognized reference samples nor reference equipment available for the determination of these explosion characteristics. Therefore the following calibration method has been carried out successfully:

A dust has been selected, prepared and supplied to **62** test laboratories all over the world. The mean values of the explosion indices, measured by the participating laboratories, has been calculated as reference values. The testing laboratories have been informed about the evaluation with a certificate.

This report presents the results of this calibration method and describes the evaluation procedures. It also demonstrates that with this method it was possible to discover and rectify the cause of any errors with installations producing results differing widely from the reference values.

### CaRo 15 - Reference values for the Explosion Indices P<sub>max</sub> and K<sub>max</sub>

<b>P<sub>max</sub></b> (bar)	<b>8.2 ± 10%</b> (7.3 ... 9.0)
<b>K<sub>max</sub></b> (bar·m/s)	<b>245 ± 10%</b> (220 ... 269)

### CaRo 15 – Reference values for the Minimum Ignition Energy MIE

<b>Es / 3</b>	<b>Es</b>	<b>Es · 3</b>
<b>0.6 mJ</b>	<b>1.7 mJ</b>	<b>5.1 mJ</b>



Birsfelden, January 2016

Adolf Kühner AG  
Christoph Cesana

## 1.1 Participants

For details see section "list of participants".

	Pmax, Kmax (75)		MIE (56)	
	20-l	1 m <sup>3</sup>	MIKE	Others
Australia	1		1	
Austria	3		2	
Belgium	4		2	
Bosnia	1			
Brazil	1			
China	1		1	
Czech Republic	1		1	
England	3		3	
France	5		3	
Germany	18	3	14	
Hungary	1			
India			2	
Italy	2		2	
Japan	1		3	
Norway	1		1	
Poland	1			
South Africa	1			
Spain	2		1	
Switzerland	4		5	
Taiwan	1		1	
U.S.A	20		12	2
<b>Total:</b>	<b>72</b>	<b>3</b>	<b>54</b>	<b>2</b>

This calibration round robin test „CaRo 15“ has been accepted worldwide as the best and most reliable calibration method for this type of test equipment.

## 1.2 Test substance

For correct calibration the CaRo 15 test sample has been milled, homogenized and shipped in an air tight package. Therefore the sample has to be tested „as delivered“.


**CaRo 15 = Niacin USP special (Nicotinic acid)**

### Particle size:

	d 10 [ $\mu\text{m}$ ]	d 50 = median [ $\mu\text{m}$ ]	d 90 [ $\mu\text{m}$ ]
Sample 1	4.3	<b>18.7</b>	52.1
Sample 2	4.5	<b>19.2</b>	52.4
Sample 3	4.5	<b>19.0</b>	51.7
Sample 4	4.9	<b>20.6</b>	62.4


## 2. Explosion Indices Pmax, Kmax

**Pmax = 8.2 bar** ± 10% (7.3 ... 9.0) by 636 g/m<sup>3</sup>



The individual results are drawn in relation to the arithmetic mean of all results and in chronological sequence (number of certificate).

**Kmax = 245 bar·m/s ± 10% (220 ... 269) by 839 g/m<sup>3</sup>**



The individual results are drawn in relation to the arithmetic mean of all results and in chronological sequence (number of certificate).

## 2.1 Test procedure:

The method for determination of Pmax, Kmax is described in the „Manual CaRo 15“.

## 2.2 Evaluation:

The explosion indices Pmax and (dP/dt)max are defined as the mean values of the maximum values of each series. Subsequently, the explosion index Kmax is calculated from the mean value (dP/dt)max.


## 2.3 Scatter of Pmax and Kmax:

The maxima of each series must not deviate by more than **10%** of Pmax resp. Kmax. Otherwise this series must be repeated!

## 2.4 Calculation of the reference values:

First the mean values of all test results (75) has been calculated. In a 2nd step all results outside of the tolerance band are excluded prior to the subsequent calculation of the mean value. Due to the large number of participants the mean values did not change.

### 3. Minimum Ignition Energy MIE



The individual results are drawn in chronological sequence (number of certificate).

#### 3.1 Test procedure:

The method for determination of the MIE is described in the „Manual CaRo 15“.

### 3.2 Estimation of the statistical energy (Es):

The minimum ignition energy MIE lies, by definition, between two energy values:

$$E_1 < \text{MIE} < E_2$$

For the purpose of comparison between different apparatus, only one MIE value (Es) instead of the energy range (E1, E2) shall be used. This single value (Es) can be estimated by use of the probability of ignition as follows (EN 13821):

Provided that for the energy E2 a minimum of 5 dust concentrations evenly distributed are tested, the position of the MIE in the E1-E2 range can be estimated. At ignition energy E2, the number of dust concentrations with ignition, is divided by the total number of dust concentrations tested.

$$E_S = 10^{\log E_2 - \frac{I[E_2] \cdot (\log E_2 - \log E_1)}{(NI + 1) \cdot [E_2] + 1}}$$

where is:  $I[E_2]$  = number of tests with ignition at the energy E2.  
 $(NI+1)[E_2]$  = total number of tests at the energy E2.

### 3.3 Criteria for conformity:

Conformity between two equipment (a, b) is given, when the Es-values differ less than a factor of 3 (EN 13821).

$$1/3 < E_s(a) / E_s(b) < 3$$

Accordingly:

Conformity in the CaRo 15 is given, when the Es-value of each equipment differ less than a factor of 3 to the mean (Es) of all equipment:


<b>Es / 3</b>	<b>Es</b>	<b>Es • 3</b>
<b>0.6 mJ</b>	<b>1.7 mJ</b>	<b>5.1 mJ</b>



## 4. Troubles with chemical igniters

### 4.1 Simex-Igniters

Simex sent some faulty igniters with pressure oscillations. Pressure oscillations increase the turbulence and therefore the Kmax-value. Through the superimposed oscillation, the automatic evaluation is influenced or even wrong, which means the tangent has to be evaluated manually. The behaviour of the igniters is strongly variable from one production lot to another. Igniters with superimposed oscillations will give false results. In that case contact the manufacturer of the igniters.



## 5. List of Participants

Country	Company Laboratory	E-Mail	Pmax Kmax	MIE
Australia	SIMTARS	fiona.clarkson@simtars.com.au	✓	✓
Austria	AUVA	silvia.springer@auva.at	✓	✓
Austria	FireX Gresslehner	dietmar.gresslehner@firex.at	✓	✓
Austria	Montanuniversität Leoben	thomas.ehgartner@unileoben.ac.at	✓	✓
Belgium	ADINEX N.V	frederik.norman@adinex.be	✓	✓
Bosnia and Herzegovina	UNIVERSITY OF TUZLA	zvjezdan.karadzin@untz.ba	✓	
Brazil	IPT-Instituto de Pesquisas	ricalca@ipt.br	✓	
China	Shanghai Research Institute of Chemical Industry	xiao_srici@163.com	✓	✓
Czech Republic	V V U U, a.s	mokosl@vvuu.cz	✓	✓
England	BRE Building Research Establishment Ltd	amendtl@bre.co.uk	✓	✓
England	Chilworth Technology Limited	dcowhig@chilworthglobal.com	✓	✓
England	Kindlow Safety Services	james.coady@kindlow.co.uk	✓	✓
England	Sigma-HSE (UK) Ltd	grogers@sigma-hse.com		✓
France	INERIS MAGASIN GL	agnes.janes@ineris.fr	✓	✓
France	Rhodia Operations	noel.michel@solvay.com	✓	✓
France	Sanofi Chimie	cecile.masson-rojas@sanofi.com	✓	✓
Germany	AQura GmbH	matthias.vorwinkel@aqura.de	✓	✓
Germany	BASF SE	Johannes.a.Fischer@basf.com	✓	✓
Germany	BAYER TECHNOLOGY SERVICES	heinz-peter.keldenich@bayer.com	✓	✓
Germany	BGN	madlen.schoenherr@bgn.de	✓	✓
Germany	BI Pharma, Boehringer Ingelheim	juergen.leininger@boehringer-ingelheim.com	✓	✓
Germany	Consilab	simon.seitz@consilab.de	✓	✓
Germany	DEKRA EXAM GmbH	matthias.reinecke@dekra.com	✓	
Germany	DGUV - IFA St. Augustin	sascha.hohmann@dguv.de	✓	✓
Germany	Henkel AG & Co. KGaA	michaela.berchter@henkel.com	✓	✓

Country	Company Laboratory	E-Mail	Pmax Kmax	MIE
Germany	IBExU Institut für Sicherheitstechnik GmbH	f.flemming@ibexu.de	✓	✓
Germany	Inburex Consulting GmbH	martin.gosewinkel@inburex.com	✓	✓
Germany	Merck KGaA	thomas.keil@merckgroup.com		✓
Germany	Siemens AG	michael.nau@siemens.com	✓	✓
Germany	Wacker Chemie AG	alfred.augsberger@wacker.com	✓	✓
Hungary	University of Miskolc	simenfalvi@uni-miskolc.hu	✓	
India	GVS Cibatech Private limited	ganesh.kumar@cibatech.com		✓
Italia	Redox SNC	info@labredox.com	✓	✓
Italia	Stazione Sperimentale	nicola.mazzei@mi.camcom.it	✓	✓
Japan	Toho KK	h.watanabe@j-toho-kk.co.jp	✓	✓
Netherlands	AKZO NOBEL Functional Chemicals B.V	albert.colomer@akzonobel.com	✓	✓
Norway	GexCon AS	jth@gexcon.com	✓	✓
Poland	Central Mining Institute	zdyduch@gig.katowice.pl	✓	
South Africa	CSIR Kloppersbos	imthombe@csir.co.za	✓	
Spain	Laboratorio Oficial J.M. Madariaga	javier.garcia@upm.es	✓	✓
Switzerland	DSM Nutritional Products AG	romeo.isner@dsm.com		✓
Switzerland	F.Hoffmann-La Roche AG	klaus.schwenzfeuer@roche.com		✓
Switzerland	Firmenich S.A.	nicolas.dubois@firmenich.com	✓	
Switzerland	Givaudan International SA	vania.glaus@givaudan.com		✓
Switzerland	Lonza	damian.venetz.vda@lonza.com	✓	
Switzerland	SWISSI Process Safety GmbH	christina.mendelin@tuev-sued.ch	✓	✓
U.S.A	Ashland Inc.	jforman@ashland.com	✓	✓
U.S.A	BASF Corporation	szu-ying.huang@basf.com	✓	✓
U.S.A	Chilworth Technology in Princeton NJ	victoria.goncalves@dekra.com	✓	
U.S.A	CYTEC Industries Inc.	mark.ryan@cytec.com	✓	✓

Country	Company Laboratory	E-Mail	Pmax Kmax	MIE
U.S.A	The Dow Chemical Company	rjbellair@dow.com	✓	✓
U.S.A	Exponent Inc.	tmyers@exponent.com	✓	
U.S.A	Fauske & Associates, LLC	kfauske@fauske.com	✓	✓
U.S.A	Fike Corporation FPS	adam.morrison@fike.com	✓	✓
U.S.A	Firmenich, Inc.	peter.de.rege@firmenich.com	✓	✓
U.S.A	FM Approvals LLC	john.travers@fmapprovals.com	✓	
U.S.A	IEP Technologies, LLC	ginger.beaupre@ieptechnologies.com	✓	
U.S.A	ioKinetic	rynkar@iokinetic.com	✓	✓
U.S.A	Merck Sharp & Dohme Corp.	michael_toth@merck.com	✓	✓
U.S.A	Texas A&M University	mashuga@tamu.edu		✓